

```
=====  
Acq. Operator   :  
Acq. Instrument : Instrument 1           Location : Vial 108  
Injection Date  : 14.02.2019 10:51:33   Inj      : 1  
                                           Inj Volume : 1 µl  
  
Acq. Method     : C:\HPCHEM\1\METHODS1  
Last changed    \SPIRT_RV.M : 14.02.2019 21:05:08  
Analysis Method : C:\CHEM32\1\METHODS\ETHANOL_AS_INTERNAL_STANDARD.M  
Last changed    : 17.05.2019 11:26:36  
Additional Info : Peak(s) manually integrated  
=====
```

Internal Standard Report

```
=====  
Sorted By       : Signal  
Calib. Data Modified : 17.05.2019 11:10:39  
Multiplier      : 1.0000  
Dilution        : 1.0000
```

Use Multiplier & Dilution Factor with ISTDs

Sample ISTD Information:

ISTD #	ISTD Amount [mg/L AA]	Name
1	7.89300e5	Ethanol

Signal 1: FID1 A,

RetTime [min]	Type	ISTD used	Area [pA*s]	Amt/Area ratio	Amount [mg/L AA]	Grp	Name
3.144	BB	1	2.01262e-1	2.13828	1.76995		Acetaldehyde
3.872		1	-	-	-		Methyl acetate
4.482		1	-	-	-		Ethyl acetate
4.651	BB	1	1.53844	1.58959	10.05774		Methanol
5.093	BV	1	1.46963e-1	1.19969	7.25122e-1		2-propanol
5.294	VB S I	1	1.91914e5	1.00000	7.89300e5		Ethanol
7.761		1	-	-	-		1-propanol
9.143		1	-	-	-		2-methyl-1-propanol
10.432		1	-	-	-		1-butanol
11.787		1	-	-	-		3-methyl-1-butanol

Totals without ISTD(s) : 12.55281

2 Warnings or Errors :

Warning : Calibration warnings (see calibration table listing)
Warning : Calibrated compound(s) not found

```
=====  
*** End of Report ***
```