

**Method OIV-MA-AS312-03A** Type II method

**Methanol**

(Resolution Oeno 377/2009, Revised by OIV-OENO 480/2014)

**1. Scope of application**

This method is applicable to the determination of methanol in wine for concentrations between 50 and 500 mg/L.

**2. Principle**

Methanol is determined in the distillate, to which ethanol is an internal standard is added, using gas chromatography with a flame ionisation detector (FID).

**3. Reagents and materials**

3.1. Type II water, according to ISO standard 3696

3.2. Ethanol: purity  $\geq 96\%$  (CAS no. 64-17-5)

3.3. Hydrogen: minimum specifications: 99.999% purity (CAS no. 1333-74-0)

3.4. Helium: minimum specifications: 99.999% purity (CAS no. 7440-59-7)

3.5. Methanol: purity  $\geq 99\%$  (CAS no. 67-56-1)

3.6. 4-Methyl-2-pentanol (internal standard): purity  $\geq 98\%$  (CAS no. 108-11-2).

Internal standard used in the validation.

*Note 1: Other internal standards can be used, such as:*

- 3-pentanol: purity  $\geq 98\%$  (CAS no. 584-02-1)
- 4-methyl-1-pentanol: purity  $\geq 98\%$  (CAS no. 626-89-1)
- Methyl nonanoate: purity  $\geq 98\%$  (CAS no. 1731-84-6)

3.7. Reference materials: these may be, for example, wines from laboratory proficiency tests.

3.8. Preparation of working solutions (by way of example):

*3.8.1. Approximately 10% v/v aqueous-alcoholic mixture*

This mixture should be as close as possible to the alcohol content of the wine to be analysed. Pour 100 mL of ethanol (3.2) into a 1 L calibrated flask (4.2), make up to volume with demineralised water (3.1) and mix.

**3.8.2. 10 g/L Internal standard solution**

Using an analytical balance (4.1), weigh approximately 1 g of internal standard (3.6) into a 100 mL calibrated flask (4.3) that contains around 60 mL of 10% ethanol solution (3.8.1), so as to minimise evaporation of the internal standard. Make up to volume with the ethanol solution (3.8.1) and mix.

**3.8.3. 1 g/L Internal standard solution**

Add 10 mL of the 10 g/L internal standard solution (3.8.2) using a pipette (4.8) and make up to 100 mL (4.3) using the 10% v/v hydroalcoholic mixture (3.8.1).

*3.8.4. 5 g/L Methanol stock solution*

Using an analytical balance (4.1), weigh approximately 500 mg of methanol (3.5) into a 100 mL calibrated flask (4.3) that contains about 60 mL of 10% ethanol solution (3.8.1), so as to minimise evaporation of the methanol. Make up to volume with the ethanol solution (3.8.1) and mix.

*3.8.5. Working calibration solutions*

By way of example, a method for plotting a calibration curve is outlined below.

Each solution should be prepared with the 10% aqueous-alcoholic mixture (3.8.1).

*3.8.5.1. 500 mg/L Methanol standard solution*

Add 10 mL of the 5 g/L stock solution (3.8.4) to a 100 mL calibrated flask (4.3) using a pipette (4.8) and make up to volume with the 10% v/v ethanol solution (3.8.1).

*3.8.5.1.1. 250 mg/L Methanol standard solution*

Add 10 mL of the 500 mg/L methanol solution (3.8.5.1) to a 20 mL calibrated flask (4.5) using a pipette (4.8) and make up to volume with the 10% v/v ethanol solution (3.8.1).

*3.8.5.1.2. 200 mg/L Methanol standard solution*

Add 20 mL of the 500 mg/L methanol solution (3.8.5.1) to a 50 mL calibrated flask (4.4) using a pipette (4.7) and make up to volume with the 10% v/v ethanol solution (3.8.1).

*3.8.5.1.3. 150 mg/L Methanol standard solution*

Add 6 mL of the 500 mg/L methanol solution (3.8.5.1) to a 20 mL calibrated flask (4.5) using a pipette (4.9) and make up to volume with the 10% v/v ethanol solution (3.8.1).

*3.8.5.1.4. 100 mg/L Methanol standard solution*

Add 4 mL of the 500 mg/L methanol solution (3.8.5.1) to a 20 mL calibrated flask (4.5) using a pipette (4.10) and make up to volume with the 10% v/v ethanol solution (3.8.1).

*3.8.5.1.5. 50 mg/L Methanol standard solution*

Add 2 mL of the 500 mg/L methanol solution (3.8.5.1) to a 20 mL calibrated flask (4.5) using a pipette (4.11) and make up to volume with the 10% v/v ethanol solution (3.8.1).

#### **4. Apparatus**

- 4.1. Analytical balance (1 mg precision)
- 4.2. 1 L Class A calibrated flasks
- 4.3. 100 mL Class A calibrated flasks
- 4.4. 50 mL Class A calibrated flasks
- 4.5. 20 mL Class A calibrated flasks
- 4.6. 10 mL Class A calibrated flasks
- 4.7. 20 mL Class A pipettes with two marks
- 4.8. 10 mL Class A pipettes with two marks
- 4.9. 6 mL Class A pipettes with two marks
- 4.10. 4 mL Class A pipettes with two marks
- 4.11. 2 mL Class A pipettes with two marks
- 4.12. 1 mL Class A pipettes with two marks or 1 mL micropipettes
- 4.13. Temperature-programmable gas chromatograph with a flame ionisation detector and a data processing system capable of calculating areas or measuring peak heights
- 4.14. Fused silica capillary column coated with a Carbowax 20M-type polar stationary phase (for example):
  - Chrompack CP-wax 57 CB, 50 m x 0.32 mm x 0.45 µm
  - DB-WAX 52, 30 m x 25 mm x 0.2 µm

#### **5. Sample preparation**

Sparkling and/or young wines must be pre-degassed, for example, by mixing 200 mL of wine in a 1 L flask. Subsequently, the samples are distilled according to the method for determining alcoholic strength by volume (OIV-MA-AS312-01A). The distillation can be carried out without adding calcium hydroxide in this case.

**5.1. Addition of internal standard (by way of example)**

Pour 10 mL of distillate into a 10 mL calibrated flask (4.6), add 1 mL (4.12) of internal standard solution (3.8.3) and mix.

**6. Procedure**

The calibration curve standards are treated in the same way as the samples (point 5.1).

It is recommended that the aqueous-alcoholic mixture (3.8.1) is injected at the start of the sequence in order to verify that it does not contain methanol.

*6.1. Operating conditions (as a guide):*

Carrier gas: helium or hydrogen

Carrier gas flow: 7 mL/min

Injection: split (ratio: 7:50)

Injection volume: 1 or 2 µL

Injector temperature: 200-260 °C

Detector temperature: 220-300 °C

Temperature programme: from 35 °C (for 2 minutes) to 170 °C, at 7.5 °C/min

**7. Calculations**

Calculate the concentration of methanol ( $C_i$ ), using the following equation:

$$C_i \text{ [mg/L AA]} = RRF_i^{\text{ethanol}} \cdot \frac{A_i}{A_{\text{ethanol}}} \cdot \rho_{\text{ethanol}}$$

$RRF_i^{\text{ethanol}}$  - relative response factor for methanol relative to ethanol

$A_i$  – peak area of methanol

$A_{\text{ethanol}}$  – peak area of ethanol

$\rho_{\text{ethanol}}$  – concentration of ethanol

## **8. Expression of the results**

The concentration of methanol is may be expressed in mg/L or in mg/100 mL absolute alcohol; in the latter case, the alcohol content by volume of the wine should be determined.

*Note 2:* mg/100 mL absolute alcohol = mg/L x 10/alcohol content by volume

## **9. Precision**

The data from the international interlaboratory test is outlined in Annex A.

## **10. Quality control**

Internal quality control may be carried out using certified reference materials or wines whose characteristics have been determined from a consensus (3.7). These should be prepared as for the samples (point 5). Participation in proficiency tests is recommended.

## **11. Report of the results**

The results are expressed to the nearest whole number (in accordance with the uncertainty).

## **12. Bibliography**

Compendium of international methods of wine and must analysis. Method OIV-MA-AS312-01A (Alcoholic strength).

## **Annex A**

### **Statistical results of the interlaboratory test**

#### **Design of validation study**

The validation study was conducted with 10 samples: 2 white wines, one dry and one sweet, 2 red wines, one of which was oaked, and 1 fortified wine (Port), including blind duplicates, according to OIV recommendations. The approximate concentration of methanol is shown in the following table.

| Sample          | White wine<br>Dry | White wine<br>Sweet | Red wine | Red wine<br>oaked | Fortified<br>wine<br>port |
|-----------------|-------------------|---------------------|----------|-------------------|---------------------------|
| Methanol (mg/L) | 50                | 150*                | 270      | 400*              | 120                       |

(\*) In this particular indicated case, methanol was added to the wine to cover a greater range of concentrations. The wine was then mixed, stabilised and bottled.

#### **Participating laboratories:**

Samples were sent to 17 laboratories in 9 different countries.

Laboratorios Agroalimentarios, Madrid (Spain)

Estación de Viticultura y Enología de Galicia, EVEGA (Spain)

Estació de Viticultura i Enologia de Vilafranca del Penedès, (Spain)

Estación Enológica de Haro, La Rioja (Spain)

Estación de Viticultura y Enología de Galicia (Spain)

Lab. Bordeaux, Service Commun des Lab., Pessac (France)

Laboratoire d'Ile-de-France, Paris (France)

Laboratoires Inter Rhône (France)

Comité Interprof. du Vin de Champagne (CIVC) (France)

Bfr-Bundesinst. f. Risikobewertung (Germany)

Landesuntersuchungsamt Mainz (Germany)

Instituto Nacional de Vitivinicultura, Mendoza (Argentina)

ALKO Inc., Alcohol Control Lab. (ACL) (Finland)

Instituto dos Vinhos do Douro e do Porto (Portugal)

Czech Agriculture and Food Inspection Authority (CAFIA), Brno (Czech Republic) CZ

National Food Safety Office, Directorate of Oenology and Alcoholic Beverages (NÉBIH BAI),  
Budapest (Hungary)

Lehr- und Forschungszentrum, Klosterneuburg (Austria)

Collaborative study on methanol

|                 | Dry white    |              | Sweet white   |               | Red    |        | Oaked red     |               | Port   |        |
|-----------------|--------------|--------------|---------------|---------------|--------|--------|---------------|---------------|--------|--------|
| Laboratory code | A            | G            | B             | H             | C      | I      | D             | J             | E      | K      |
| A               | 39.99        | 38.13        | 127.42        | 136.25        | 144.80 | 145.71 | 496.53        | 513.00        | 192.13 | 219.39 |
| B               | 41.20        | 40.90        | 157.60        | 160.50        | 150.40 | 146.90 | 484.90        | 477.80        | 222.40 | 219.60 |
| C               | 36.80        | 35.60        | 133.50        | 129.20        | 119.10 | 134.10 | 454.10        | 478.40        | 197.00 | 174.80 |
| D               | 36.00        | 39.60        | <b>177.40</b> | <b>145.50</b> | 160.80 | 138.00 | <b>302.00</b> | <b>494.50</b> | 216.10 | 248.50 |
| E               | <b>68.00</b> | <b>70.00</b> | 163.00        | 169.00        | 178.00 | 177.00 | 503.00        | 495.50        | 225.00 | 227.00 |
| F               | 37.00        | 37.10        | 148.30        | 148.20        | 143.40 | 142.40 | 484.10        | 474.00        | 206.30 | 206.90 |
| G               | 41.40        | 42.30        | 152.60        | 152.40        | 149.70 | 150.50 | 489.60        | 491.10        | 216.60 | 217.20 |
| H               | 36.80        | 32.40        | 140.80        | 129.10        | 128.00 | 137.70 | 440.60        | 429.30        | 187.50 | 192.80 |
| I               | 42.90        | 43.30        | 153.50        | 155.50        | 139.70 | 147.40 | 468.30        | 456.10        | 225.30 | 225.60 |
| J               | 40.90        | 40.60        | 155.50        | 154.60        | 148.50 | 149.10 | 496.40        | 499.80        | 217.10 | 217.00 |
| K               | 39.30        | 36.20        | <b>103.10</b> | <b>143.10</b> | 131.90 | 115.90 | <b>437.90</b> | <b>334.00</b> | 156.10 | 172.60 |
| L               | 35.00        | 39.00        | 164.00        | 167.00        | 157.00 | 160.00 | 492.00        | 508.00        | 249.00 | 220.00 |
| M               | 43.60        | 43.40        | 157.30        | 154.90        | 155.50 | 158.90 | 506.80        | 496.10        | 217.70 | 219.50 |
| N               | 34.20        | 33.60        | 126.50        | 125.70        | 125.90 | 133.60 | 429.10        | 429.00        | 192.10 | 188.90 |
| O               | 34.00        | 35.70        | 149.00        | 154.80        | 144.20 | 141.80 | 482.80        | 473.60        | 210.40 | 218.10 |
| P               | 44.70        | 43.70        | 151.60        | 146.90        | 140.70 | 147.60 | 451.20        | 472.80        | 205.40 | 205.80 |
| Q               | 40.70        | 38.80        | 153.00        | 149.80        | 158.20 | 153.40 | 498.20        | 497.50        | 225.50 | 217.20 |

Note: The values in bold correspond to values rejected according to the Cochran (variance outliers) and Grubbs (mean outliers) tests.

| Indicators                         | Dry white | Sweet white | Red       | Oaked red | Port      |
|------------------------------------|-----------|-------------|-----------|-----------|-----------|
| Number of accepted laboratories    | 16        | 15          | 17        | 15        | 17        |
| Number of repetitions              | 2         | 2           | 2         | 2         | 2         |
| Minimum                            | 32.40     | 125.70      | 115.90    | 429.00    | 156.10    |
| Maximum                            | 44.70     | 169.00      | 178.00    | 513.00    | 249.00    |
| Repeatability variance $s_r^2$     | 2.2466    | 12.1330     | 39.0164   | 76.3567   | 105.3390  |
| Intergroup variance $s_r^2$        | 9.61893   | 146.39249   | 151.90249 | 535.61827 | 292.14282 |
| Reproducibility variance $s_r^2$   | 11.8655   | 158.5254    | 190.9189  | 611.9750  | 397.4819  |
| Overall mean                       | 38.90     | 148.92      | 145.76    | 478.97    | 210.37    |
| Repeatability standard deviation   | 1.50      | 3.48        | 6.25      | 8.74      | 10.26     |
| r Limit                            | 4.242     | 9.858       | 17.677    | 24.729    | 29.046    |
| Repeatability CV                   | 3.9       | 2.3         | 4.3       | 1.8       | 4.9       |
| Reproducibility standard deviation | 3.44      | 12.59       | 13.82     | 24.74     | 19.94     |
| R Limit                            | 9.748     | 35.632      | 39.103    | 70.009    | 56.422    |
| Reproducibility CV                 | 8.9       | 8.5         | 9.5       | 5.2       | 9.5       |
| Horwitz RSD                        | 6.09      | 4.97        | 4.99      | 4.17      | 4.72      |
| Horrat r                           | 0.6       | 0.5         | 0.9       | 0.4       | 1.0       |
| Horwitz RSD                        | 9.22      | 7.53        | 7.56      | 6.32      | 7.15      |
| Horrat R                           | 1.0       | 1.1         | 1.3       | 0.8       | 1.3       |

According to the Horrat values, the repeatability and reproducibility of the method are acceptable

**Z-scores obtained by the participants: of the 85 Z-scores, 3 are unsatisfactory and 4 are questionable**

|                 | Z-score        | Z-score          | Z-score     | Z-score        | Z-score      |
|-----------------|----------------|------------------|-------------|----------------|--------------|
| Laboratory code | Dry white wine | Sweet white wine | Red wine    | Oaked red wine | Port         |
| A               | 0.05           | -1.36            | -0.04       | 1.04           | -0.23        |
| B               | 0.62           | 0.80             | 0.21        | 0.10           | 0.53         |
| C               | -0.78          | -1.40            | -1.39       | -0.51          | -1.23        |
| D               | -0.32          | 1.00             | 0.26        | <b>-3.26</b>   | 1.10         |
| E               | <b>8.74</b>    | 1.36             | <b>2.30</b> | 0.81           | 0.78         |
| F               | -0.54          | -0.05            | -0.21       | 0.00           | -0.19        |
| G               | 0.86           | 0.28             | 0.31        | 0.46           | 0.33         |
| H               | -1.25          | -1.11            | -0.93       | -1.78          | -1.01        |
| I               | 1.22           | 0.44             | -0.16       | -0.68          | 0.76         |
| J               | 0.54           | 0.49             | 0.22        | 0.77           | 0.34         |
| K               | -0.33          | <b>-2.05</b>     | -1.58       | <b>-3.76</b>   | <b>-2.31</b> |
| L               | -0.55          | 1.32             | 0.92        | 0.85           | 1.21         |
| M               | 1.34           | 0.57             | 0.83        | 0.91           | 0.41         |
| N               | -1.45          | -1.81            | -1.16       | <b>-2.02</b>   | -1.00        |
| O               | -1.18          | 0.24             | -0.20       | -0.03          | 0.19         |
| P               | 1.54           | 0.03             | -0.12       | -0.69          | -0.24        |
| Q               | 0.25           | 0.20             | 0.73        | 0.76           | 0.55         |