



Calculation of Volatile Compound Concentrations in Alcoholic Beverages

The calculator is designed for comparative evaluation of the metrological characteristics of various chromatographic methods for analysis of alcoholic and other alcohol-containing products

PARAMETERS OF METHODS

SS CONCENTRATIONS

SS AREAS

CALIBRATION

SAMPLES OF PRODUCTS

RE

Basic parameters of methods and standard solutions (SS)

1. The following chromatographic methods are used to analyze alcohol-containing products: external standard method using a set of standard solutions (SS), traditional internal standard method using an additional substance added to the SS and product samples at a known concentration, and new modified internal standard method (IS_Ethanol), in which ethanol contained in the SS and products is used as an internal standard.
2. The initial stage of all methods is calibration — the determination of calibration characteristics by the analysis of SS with certified concentration values C^{cert} of volatile compounds. Then, using the established calibration parameters, the content of analyzed compounds in samples C^{meas} is determined.
3. All methods use multiple SS containing compounds in known concentrations. Each set of SS corresponds to several concentration levels, which are needed for determination of the calibration coefficients and verification the linearity of the detector response. It is necessary to specify the count of analyte compounds and the count of SS to work of the calculator.
4. Different methods of alcohol analysis may use different units of concentration measurement. To ensure accurate calculations and easy comparison of different methods, concentration values are converted from "primary" units of measurement to "final" units, which will be used in the calculations in the methods. The calculator offers the following unit options (abbreviation AA - anhydrous alcohol): mg/L ($mg/L = mg/dm^3$), $mg/L AA$, $vol.\%$, $vol.\% AA$, $g/100L AA$ ($g/100L AA = mg/100cm^3 AA$), $\mu g/g$ ($\mu g/g = ug/g$).
5. The IS_Ethanol method calculates per AA, so the following can be set as "final" units of measurement for IS_Ethanol: $mg/L AA$, $vol.\% AA$, $g/100L AA$, $mg/100cm^3 AA$. If the "primary" units of measurement are also per AA, then information about the strength of SS and product samples is not required.
6. To calibrate a chromatograph using methods that use an internal standard (traditional method or IS_Ethanol), the relative response factors (RRFs) are determined. A single calibration level is used as a rule with an "intermediate" concentration values. The remaining levels are used to evaluate the linearity of the detector response.

Enter the count of compounds

Count of compounds:

3

Enter the count of concentration levels

Count of concentration levels:

3

Enter the number of calibration level (for RRF calculation)

Calibration level number:

3

Enter the primary units of concentration ($\mu\text{g/g} = \text{ug/g}$)

Primary units of concentration:

ug/č ✓

Enter the final concentration units in which the results should be calculated.

Final units of concentration:

vol.% ✓

When converting units of measurement per anhydrous alcohol, enter the SS strength value, in vol.% (from 0.5 to 100)

Strength of SS, vol.%:

40

If the units of measurement are $\mu\text{g/g}$ ($\mu\text{g/g} = \text{ug/g}$), enter the value of SS density at a temperature of 20 °C, in $\text{kg/m}^3 = \text{g/l}$ (from 772 to 1000)

Density of SS, kg/m^3 :

948.06

To button "Submit values" ->

Back to Main Menu ->

Concentrations of compounds in standard solutions (SS). Conversion of concentration units

Fill in the table:

Table 1. Concentration values of compounds in SS, expressed in primary units $\mu\text{g/g}$

| № | Compound | Concentration level | | |
|---|--------------|---------------------|--------|--------|
| | | 1 | 2 | 3 |
| 1 | methanol ✓ | 21.634 | 107.54 | 203.11 |
| 2 | ethanol ✓ | 0 | 0 | 0 |
| 3 | 3-pentanol ✓ | 228.73 | 227.65 | 226.16 |

To button "Submit values" ->

Back to Main Menu ->

Table 2. Concentration values of compounds in SS, converted into specified units of measurement vol.% AA

| № | Compound | Concentration level | | |
|---|------------|---------------------|-------------------|-------------------|
| | | 1 | 2 | 3 |
| 1 | methanol | 0.0064758556579944 | 0.032190696009093 | 0.060798328681485 |
| 2 | ethanol | 100 | 100 | 100 |
| 3 | 3-pentanol | 0.066518332453988 | 0.066204251226994 | 0.065770935460123 |

Data from chromatograms of SS

Enter the count of repeated measurements for the concentration level 1

Count of repeated measurements

2

Concentration level 1.

Peak area values of compounds

| № | Compound | Concentration, vol.% AA | Pick area, units of area | |
|---|------------|-------------------------|--------------------------|----------|
| | | | 1 | 2 |
| 1 | methanol | 0.0064758556579944 | 0.00164 | 0.001655 |
| 2 | ethanol | 100 | 29.0017 | 28.957 |
| 3 | 3-pentanol | 0.066518332453988 | 0.0346 | 0.03461 |

To button "Submit values" ->

[Back to Main Menu](#) ->

Enter the count of repeated measurements for the concentration level 2

Count of repeated measurements

2

Concentration level 2.

Peak area values of compounds

| № | Compound | Concentration, vol.% AA | Pick area, units of area | |
|---|------------|-------------------------|--------------------------|----------|
| | | | 1 | 2 |
| 1 | methanol | 0.032190696009093 | 0.00805 | 0.00796 |
| 2 | ethanol | 100 | 28.9454 | 28.94586 |
| 3 | 3-pentanol | 0.066204251226994 | 0.015244 | 0.015139 |

To button "Submit values" ->

[Back to Main Menu](#) ->

Enter the count of repeated measurements for the concentration level 3

Count of repeated measurements

2

Calibration level for *RRF*

Concentration level 3.

Peak area values of compounds

| № | Compound | Concentration, vol.% AA | Pick area, units of area | |
|---|------------|----------------------------|--------------------------|---------|
| | | | 1 | 2 |
| 1 | methanol | 0.060798328681485 | 0.0151 | 0.01512 |
| 2 | ethanol | 100 | 28.808 | 29.0155 |
| 3 | 3-pentanol | 0.065770935460123 | 0.0341 | 0.0343 |

To button "Submit values" ->

Back to Main Menu ->

Calibration

Select calibration methods:

- modified internal standard method with ethanol
- external standard method
- internal standard method with additional reference

If you select internal standard method, please indicate reference substance

Reference substance (internal standard):

3-pentanol



To button "Submit values" ->

Back to Main Menu ->

Calibration by modified internal standard method (reference - ethanol) ->

Calibration by external standard method ->

Calibration by internal standard method (reference - 3-pentanol) ->

Calibration by modified internal standard method (reference - ethanol)

Calibration $C/C_{eth} = RRF \cdot A/A_{eth}$; $RRF = 1.1633041442643$

Compound №1: methanol

| C_{cert} , vol.% AA | C_{meas} , vol.% AA | SD, vol.% AA | Repeatability limit, vol.% AA | RSD, % | Bias, % | Count of repeats |
|--------------------------|--------------------------|-----------------|-------------------------------------|-----------|------------|---------------------|
| | | | | | | |

| | | | | | | |
|---------|---------|---------|---------|------|-------|---|
| 0.00648 | 0.00661 | 0.00005 | 0.00014 | 0.75 | 2.13 | 2 |
| 0.03219 | 0.03217 | 0.00026 | 0.00071 | 0.80 | -0.06 | 2 |
| 0.06080 | 0.06080 | 0.00025 | 0.00070 | 0.41 | -0.00 | 2 |

Calibration $C/C_{eth} = RRF \cdot A/A_{eth}$; $RRF = 0.55600803184971$

Compound №3: 3-pentanol

| C_{cert} , vol.% AA | C_{meas} , vol.% AA | SD, vol.% AA | Repeatability limit, vol.% AA | RSD, % | Bias, % | Count of repeats |
|--------------------------|--------------------------|-----------------|-------------------------------------|-----------|------------|---------------------|
| 0.06652 | 0.06639 | 0.00009 | 0.00024 | 0.13 | -0.19 | 2 |
| 0.06620 | 0.02918 | 0.00014 | 0.00040 | 0.49 | -55.92 | 2 |
| 0.06577 | 0.06577 | 0.00006 | 0.00017 | 0.09 | -0.00 | 2 |

Calibration $C/C_{eth} = RRF \cdot A/A_{eth}$

Coefficients of calibration and FID response linearity

| Nº | Compound | Coefficient of calibration $RRF_{ethanol}$ | Coefficient of determination R^2 | Coefficient of correlation C_P |
|----|------------|--|--|--|
| 1 | methanol | 1.16 | 0.99984 | 0.99995 |
| 2 | ethanol | 1 | - | - |
| 3 | 3-pentanol | 0.556 | 0.01212 | 0.99999 |

To button "Submit values" ->

[Back to Main Menu](#) ->

☞ Calibration by modified internal standard method (reference - ethanol) ->

☞ Calibration by external standard method ->

☞ Calibration by internal standard method (reference - 3-pentanol) ->

Calibration by external standard method

Calibration $C = RF \cdot A$; $RF = 4.0223104001822$ (vol.% AA)/(area units)

Compound №1: methanol

| C_{cert} , vol.% AA | C_{meas} , vol.% AA | SD, vol.% AA | Repeatability limit, vol.% AA | RSD, % | Bias, % | Count of repeats |
|--------------------------|--------------------------|-----------------|-------------------------------------|-----------|------------|---------------------|
| 0.00648 | 0.00663 | 0.00004 | 0.00012 | 0.64 | 2.33 | 2 |
| 0.03219 | 0.03220 | 0.00026 | 0.00071 | 0.79 | 0.02 | 2 |
| 0.06080 | 0.06078 | 0.00006 | 0.00016 | 0.09 | -0.03 | 2 |

Calibration $C = RF \cdot A$; $RF = 2.1389923348166$ (vol.% AA)/(area units)

Compound №3: 3-pentanol

| $C_{\text{cert}},$ vol.% AA | $C_{\text{meas}},$ vol.% AA | SD, vol.% AA | Repeatability limit, vol.% AA | RSD, % | Bias, % | Count of repeats |
|--------------------------------|--------------------------------|-----------------|-------------------------------------|-----------|------------|---------------------|
| 0.06652 | 0.07402 | 0.00002 | 0.00004 | 0.02 | 11.28 | 2 |
| 0.06620 | 0.03249 | 0.00016 | 0.00044 | 0.49 | -50.92 | 2 |
| 0.06577 | 0.07315 | 0.00030 | 0.00084 | 0.41 | 11.22 | 2 |

Calibration $C = RF \cdot A$

Coefficients of calibration and FID response linearity

| № | Compound | Coefficient of calibration $RF,$ (vol.% AA)/(area un.) | Coefficient of determination R^2 | Coefficient of correlation C_P |
|---|------------|---|--|--|
| | | | | |
| 1 | methanol | 4.02 | 0.99996 | 0.99999 |
| 2 | ethanol | 0 | - | - |
| 3 | 3-pentanol | 2.14 | -0.00231 | -0.07352 |

To button "Submit values" ->

Back to Main Menu ->

☞ Calibration by modified internal standard method (reference - ethanol) ->

☞ Calibration by external standard method ->

☞ Calibration by internal standard method (reference - 3-pentanol) ->

Calibration by internal standard method (reference - 3-pentanol)

Calibration $C/C_{\text{st}} = RRF \cdot A/A_{\text{st}}$; $RRF = 2.092252852374$

Compound №1: methanol

| $C_{\text{cert}},$ vol.% AA | $C_{\text{meas}},$ vol.% AA | SD, vol.% AA | Repeatability limit, vol.% AA | RSD, % | Bias, % | Count of repeats |
|--------------------------------|--------------------------------|-----------------|-------------------------------------|-----------|------------|---------------------|
| 0.00648 | 0.00663 | 0.00004 | 0.00011 | 0.62 | 2.32 | 2 |
| 0.03219 | 0.07299 | 0.00022 | 0.00062 | 0.31 | 126.74 | 2 |
| 0.06080 | 0.06080 | 0.00019 | 0.00054 | 0.32 | -0.00 | 2 |

Calibration $C/C_{\text{st}} = RRF \cdot A/A_{\text{st}}$

Coefficients of calibration and FID response linearity

| № | Compound | Coefficient of calibration $RRF_{3\text{-pentanol}}$ | Coefficient of determination R^2 | Coefficient of correlation C_P |
|---|------------|--|--|--|
| | | | | |
| 1 | methanol | 2.09 | 0.97194 | 0.99999 |
| 2 | ethanol | 1.8 | - | - |
| 3 | 3-pentanol | 1 | - | - |

To button "Submit values" ->

[Back to Main Menu ->](#)

✓ Calibration by modified internal standard method (reference - ethanol) ->

✓ Calibration by external standard method ->

✓ Calibration by internal standard method (reference - 3-pentanol) ->

Parameters of the samples of the products

Enter the count of the samples of the alcoholic beverages

Count of samples:

[To button "Submit values" ->](#)

[Back to Main Menu ->](#)

Enter a name of the product sample and the count of its repeated measurements

Sample's name:

Count of repeated measurements:

To perform calculations per anhydrous alcohol by external standard method, indicate the strength value of the product sample

Strength of the product sample, vol.% (from 0.5 to 100):

If you use units $\mu\text{g/g}$ and external standard method, provide the density value of the product sample

Density of the product sample at temperature of 20 $^{\circ}\text{C}$, in $\text{kg/m}^3 = \text{g/l}$ (from 772 to 1000):

If you need to perform calculation by internal standard method, then indicate the value of concentration of the internal standard substance in the product sample

Concentration of substance "3-pentanol" in the product, in $\mu\text{g/g}$:

[To button "Submit values" ->](#)

[Back to Main Menu ->](#)

Peak area values of compounds in sample "Wine"

| № | Compound | Pick area, units of area | |
|---|------------|--------------------------|----------|
| | | 1 | 2 |
| 1 | methanol | 0.01695 | 0.01704 |
| 2 | ethanol | 12.9809 | 13.04125 |
| 3 | 3-pentanol | 0.03306 | 0.033353 |

[To button "Submit values" ->](#)

[Back to Main Menu ->](#)

Results for sample "Wine", obtained by modified internal standard method (reference - ethanol)

Calibration $C/C_{eth} = RRF \cdot A/A_{eth}$

| No | Compound | C^{meas} , vol.% AA | SD, vol.% AA | Repeatability limit, vol.% AA | RSD, % | $RRF^{ethanol}$ |
|----|------------|--------------------------|-----------------|-------------------------------------|-----------|-----------------|
| 1 | methanol | 0.15195 | 0.00007 | 0.00020 | 0.05 | 1.16 |
| 2 | 3-pentanol | 0.14190 | 0.00042 | 0.00116 | 0.30 | 0.556 |

[Back to Main Menu ->](#)

🔗 Results using by the internal standard "ethanol" ->

🔗 Results using by external standard method->

🔗 Results using by the internal standard "3-pentanol" ->

Results for sample "Wine", obtained by external standard method

Calibration $C = RF \cdot A$, (vol.% AA)/(area units)

| No | Compound | C^{meas} , vol.% AA | SD, vol.% AA | Repeatability limit, vol.% AA | RSD, % | RF , (vol.% AA)/(area un.) |
|----|------------|--------------------------|-----------------|-------------------------------------|-----------|---------------------------------|
| 1 | methanol | 0.15107 | 0.00057 | 0.00157 | 0.37 | 4.02 |
| 2 | 3-pentanol | 0.15697 | 0.00098 | 0.00271 | 0.62 | 2.14 |

[Back to Main Menu ->](#)

🔗 Results using by the internal standard "ethanol" ->

🔗 Results using by external standard method ->

🔗 Results using by the internal standard "3-pentanol" ->

Results for sample "Wine", obtained by internal standard method (reference - 3-pentanol)

Calibration $C/C_{st} = RRF \cdot A/A_{st}$

| No | Compound | C^{meas} , vol.% AA | SD, vol.% AA | Repeatability limit, vol.% AA | RSD, % | $RRF^{3-pentanol}$ |
|----|------------|--------------------------|-----------------|-------------------------------------|-----------|--------------------|
| 1 | methanol | 0.15163 | 0.00038 | 0.00105 | 0.25 | 2.09 |
| 2 | 3-pentanol | 0.14160 | 0.00000 | 0.00000 | 0.00 | 1 |

[Back to Main Menu ->](#)

- Results using by the internal standard "ethanol" ->
- Results using by external standard method ->
- Results using by the internal standard "3-pentanol" ->

Enter a name of the product sample and the count of its repeated measurements

Sample's name:

Count of repeated measurements:

To perform calculations per anhydrous alcohol by external standard method, indicate the strength value of the product sample

Strength of the product sample, vol.% (from 0.5 to 100):

If you use units $\mu\text{g/g}$ and external standard method, provide the density value of the product sample

Density of the product sample at temperature of 20 °C, in $\text{kg/m}^3 = \text{g/l}$ (from 772 to 1000):

If you need to perform calculation by internal standard method, then indicate the value of concentration of the internal standard substance in the product sample

Concentration of substance "3-pentanol" in the product, in $\mu\text{g/g}$:

[To button "Submit values" ->](#)

[Back to Main Menu ->](#)

Peak area values of compounds in sample "Whiskey"

| № | Compound | Pick area, units of area | |
|---|------------|--------------------------|-----------|
| | | 1 | 2 |
| 1 | methanol | 0.002258 | 0.002175 |
| 2 | ethanol | 29.5443 | 29.4638 |
| 3 | 3-pentanol | 0.03689 | 0.0339845 |

[To button "Submit values" ->](#)

[Back to Main Menu ->](#)

[To the next product ->](#)

Results for sample "Whiskey", obtained by modified internal standard method (reference - ethanol)

Calibration $C/C_{eth} = RRF \cdot A/A_{eth}$

| No | Compound | C ^{meas} , vol.% AA | SD, vol.% AA | Repeatability limit, vol.% AA | RSD, % | RRF ^{ethanol} |
|----|------------|---------------------------------|-----------------|-------------------------------------|-----------|------------------------|
| 1 | methanol | 0.00874 | 0.00021 | 0.00059 | 2.45 | 1.16 |
| 2 | 3-pentanol | 0.06678 | 0.00374 | 0.01037 | 5.60 | 0.556 |

[Back to Main Menu ->](#)

🔗 Results using by the internal standard "ethanol" ->

🔗 Results using by external standard method->

🔗 Results using by the internal standard "3-pentanol" ->

Results for sample "Whiskey", obtained by external standard method

Calibration $C = RF \cdot A$, (vol.% AA)/(area units)

| No | Compound | C ^{meas} , vol.% AA | SD, vol.% AA | Repeatability limit, vol.% AA | RSD, % | RF, (vol.% AA)/(area un.) |
|----|------------|---------------------------------|-----------------|-------------------------------------|-----------|------------------------------|
| 1 | methanol | 0.00870 | 0.00023 | 0.00064 | 2.65 | 4.02 |
| 2 | 3-pentanol | 0.07395 | 0.00429 | 0.01188 | 5.80 | 2.14 |

[Back to Main Menu ->](#)

🔗 Results using by the internal standard "ethanol" ->

🔗 Results using by external standard method ->

🔗 Results using by the internal standard "3-pentanol" ->

Results for sample "Whiskey", obtained by internal standard method (reference - 3-pentanol)

Calibration $C/C_{st} = RRF \cdot A/A_{st}$

| No | Compound | C ^{meas} , vol.% AA | SD, vol.% AA | Repeatability limit, vol.% AA | RSD, % | RRF ^{3-pentanol} |
|----|------------|---------------------------------|-----------------|-------------------------------------|-----------|---------------------------|
| 1 | methanol | 0.00833 | 0.00026 | 0.00073 | 3.15 | 2.09 |
| 2 | 3-pentanol | 0.06357 | 0.00000 | 0.00000 | 0.00 | 1 |

[Back to Main Menu ->](#)

🔗 Results using by the internal standard "ethanol" ->

🔗 Results using by external standard method ->

🔗 Results using by the internal standard "3-pentanol" ->

PRESS TO SUBMIT DATA